

Shoulder P.A.D. II— **Pillow Abduction Device**

Single Patient Use Only • Rx Only

Small 11630005 11630006 Medium 11630007 Large

IMPORTANT INFORMATION

Please read all instructions before use. Correct application is essential to proper functioning of product.

INDICATIONS

Soft tissue repairs. Shoulder sprains and strains. Acute shoulder instability, Surgical rotator cuff repair, Anterior glenohumeral capsular joint ligament reconstructions, Acute rotator cuff tendinopathy, S/P glenohumeral acromioplasty, Dislocation/subluxation of G-H ioint. Acromioclavicular joint dislocations, Bankart repairs

PRECAUTIONS

 This product is to be fitted initially by a physician (or properly licensed practitioner) who is familiar with the purpose for which they are responsible. The physician or practitioner is responsible for providing wearing instructions and precautions to other healthcare practitioners, the patient, or other care providers involved in the patient's care.

· Consult your physician (or properly licensed practitioner) immediately if you experience sensation changes, unusual reactions, swelling or prolonged pain while using this product. · Follow the instructions of your physician (or properly licensed practitioner) for length and duration of use.



INSTRUCTIONS FOR USE

1. Detach shoulder strap from rings at front of

arm sling. Place affected arm in sling so that it rests in the fold of the sling (1a) (Rest thumb on adjustable strap for additional comfort if desired) (1b). Attach Velcro® strap across top of sling for closed pouch effect (optional) (1c).

2. Bring strap across the back and opposite shoulder (2a) and pull through both rings at front of sling. Secure Velcro[®] end back onto

sling strap (2b). Note: Depending on right or left fitting, adjust Y-tab Velcro[®] strap ends so that the strap(s)

secures back onto itself properly. Strap length may also be adjusted by removing Y-tab end, trimming the strap, and reapplying Y-tab end.













DeRoyal

Shoulder P.A.D. II— **Pillow Abduction Device**

Single Patient Use Only • Rx Only

Small 11630005 11630006 Medium 11630007 Large

IMPORTANT INFORMATION

Please read all instructions before use. Correct application is essential to proper functioning of product.

INDICATIONS

Soft tissue repairs. Shoulder sprains and strains. Acute shoulder instability, Surgical rotator cuff repair, Anterior glenohumeral capsular joint ligament reconstructions. Acute rotator cuff tendinopathy, S/P glenohumeral acromioplasty, Dislocation/subluxation of G-H joint, Acromioclavicular joint dislocations, Bankart repairs

PRECAUTIONS

 This product is to be fitted initially by a physician (or properly licensed practitioner) who is familiar with the purpose for which they are responsible. The physician or practitioner is responsible for providing wearing instructions and precautions to other healthcare practitioners, the patient, or other care providers involved in the patient's care.

· Consult your physician (or properly licensed practitioner) immediately if you experience sensation changes, unusual reactions, swelling or prolonged pain while using this product.

· Follow the instructions of your physician (or properly licensed practitioner) for length and duration of use.



INSTRUCTIONS FOR USE

1. Detach shoulder strap from rings at front of

arm sling. Place affected arm in sling so that it rests in the fold of the sling (1a) (Rest thumb on adjustable strap for additional comfort if desired) (1b) Attach Velcro® strap across top of sling for closed pouch effect (optional) (1c). 2. Bring strap across the

back and opposite

shoulder (2a) and pull

through both rings at

front of sling. Secure

Velcro[®] end back onto sling strap (2b). Note:

Depending on right or

left fitting, adjust Y-tab

Velcro[®] strap ends so

properly. Strap length

may also be adjusted

by removing Y-tab end,

trimming the strap, and

reapplying Y-tab end.


that the strap(s)















- Align foam pillow at waistline on affected side (3a). Attach arm sling to pillow by securing hook and loop portions of Velcro® together (3b).
- 4. Detach clasp at end of pillow and place waist strap around patient (4a). Adjust length as necessary before buckling clasp back into the receiving end (4b). Pull waist strap tightly before securing Y-tab Velcro® end back onto strap (4c). The position of the pillow may be adjusted by sliding the pillow



sliding the pillow forward or back along the waistline to achieve the desired external or internal rotation. If

pillow can be pushed downward with pressure, the waist strap should be tightened. Stress ball included to promote proper circulation and encourage exercise.

CARE INSTRUCTIONS

Hand wash arm sling and pillow cover in luke warm water and air dry.



Manufacturer DeRoyal 200 DeBusk Lane, Powell, TN 37849 U.S.A. 888.938.7828 865.938.7828

Authorized Representative DeRoyal Europe, Ltd. Virginia Road, Kells, County Meath Ireland 353.46.9280100

Part #0-1745 Revised 8/03 ©2003 DeRoyal. All Rights Reserved.



- Align foam pillow at waistline on affected side (3a). Attach arm sling to pillow by securing hook and loop portions of Velcro® together (3b).
- Detach clasp at end of pillow and place waist strap around patient (4a). Adjust length as necessary before buckling clasp back into the receiving end (4b). Pull waist strap

tightly before securing Y-tab Velcro[®] end back onto strap (4c). The position of the pillow

may be adjusted by sliding the pillow forward or back along the waistline to achieve

the desired external or internal rotation. If pillow can be pushed downward with pressure, the waist strap should be tightened. Stress ball included to promote proper circulation and encourage exercise.

CARE INSTRUCTIONS

Hand wash arm sling and pillow cover in luke warm water and air dry.



Manufacturer DeRoyal 200 DeBusk Lane, Powell, TN 37849 U.S.A. 888.938.7828 865.938.7828

Authorized Representative

DeRoyal Europe, Ltd. Virginia Road, Kells, County Meath Ireland 353.46.9280100

Part #0-1745 Revised 8/03 ©2003 DeRoyal. All Rights Reserved.